

We, Atos Medical AB, hereby declare that the below mentioned devices comply with Swedish regulation LVFS 2003:11, transposing Annex II, with the exemption of section 4 (for class IIa and class IIb devices) and Annex VII (for class I devices) of the European Medical Devices Directive 93/42/EEC and clause 3.5 of Schedule 3 to the Therapeutic Goods (Medical Devices) Regulations 2002.

The Trach-HME Products

| REF | Name | Class | GMDN code |
|------|---------------------------|-------|-----------|
| 7704 | TrachPhone (50 pcs) | IIa | 58705 |
| 7705 | MEDIFLUX HCH F6 (Medival) | IIa | 58705 |
| 7707 | TrachPhone (30 pcs) | IIa | 58705 |
| 7723 | TrachPhone (5 pcs) | IIa | 58705 |

Each kind of medical device to which the system has been applied complies with the applicable provisions of the Essential Principles and Essential Requirements, the classification rules, applicable standards and the full quality assurance procedures, at each stage, from the design of the device until its final inspection before being supplied.

Valid Certificates are controlled and filed by QA/RA Department. For standards applied see Applicable Standards & Essential Requirements for each product.

Notified Body: Intertek Semko AB, Sweden. Identification no. 0413
EC-certificate no. 41310296-04

Competent Authority: Medical Products Agency, Sweden

Atos Medical AB
Kraftgatan 8
SE-242 35 Hörby, Sweden
Tel: +46 (0) 415 198 00
E-mail: info@atosmedical.com

Document Approvals
Approved Date: 2026-05-20

| | |
|-----------------------------------|---|
| Approval Task Verdict: Approve | MALIN.NYSTROM Malin Nyström, Consultant (malin.nystrom-atosmedical@coloplast.com) Issuer 20-May-2026 10:59:34 GMT+0000 |
| Approval Task Verdict: Approve | HENRIK.HERINGSLACK Henrik Heringslack, Site Director (henrik.heringslack-atosmedical@coloplast.com) Management 20-May-2026 11:40:38 GMT+0000 |
| Approval Task Verdict: Approve | KARNIL Karolina Nilsson, Head of RA, Voice & Respiratory Care (karolina.nilsson-atosmedical@coloplast.com) Regulatory 20-May-2026 11:47:16 GMT+0000 |